

Neotech Ocular Exam Kit


DISPOSABLE SPECULUM AND DEPRESSOR



FEATURES & BENEFITS

- · Disposable, sterile and ready to use
- Unique speculum designed to:
 - Lock in place with eyelid pressure
 - Allow unrestricted view
 - Open and close with ease
- Speculum available in two sizes
- Neonatal speculum (N6501) ideal for ROP exam
- Two-sided depressor with innovative texture intended to gently grip the conjunctiva
- Not made with natural rubber latex or plasticizer DEHP
- Individually packaged

CAT.#	ITEM	QTY/UNIT
N6501	Neonatal Speculum/Depressor	10/box
N6511	Standard Speculum/Depressor	10/box







Neotech Ocular Exam Kit

DISPOSABLE SPECULUM AND DEPRESSOR

INDICATIONS FOR USE

The Neotech Ocular Exam Kit is intended to assist in ocular examinations.

DIRECTIONS FOR USE

- 1 Move sliding arm to closed position.
- 2 Insert curved blades under margins of upper and lower eyelids.
- 3 Move sliding arm to open eyelids to desired spacing and release.
- 4 Use depressor to assist in visualization.
- 5 Slide sliding arm to closed position and remove curved blades from eyelid margins.
- 6 Discard speculum and depressor.

Bilateral Examination

Discard speculum and depressor. Repeat steps above with second speculum and depressor.

Caution: To prevent cross contamination, use one set of instruments per eye.

GENERAL INFORMATION

Product Materials

Speculum: Thermoplastic resin Depressor: Thermoplastic resin Wrap: Paper, polyethylene film

This product is not made with natural rubber

latex or plasticizer DEHP

Size

Speculum maximum aperture: 23 mm Depressor cylindrical head width: 9 mm Depressor spoon head width: 5.5 mm

Packaging

Individually packaged: Yes Package type: Pouch

Visit neotechproducts.com for the most up-to-date information.

Pouch is NXL film and reinforced coated paper.

— CAUTIONS & WARNINGS —

Federal law restricts this device to sale by or on the order of a physician.

The Neotech Ocular Exam Kit should only be used while the patient is under the continuous, direct supervision of healthcare professionals.

To prevent cross contamination, use one set of instruments per eye.

Excessive pressure to the eye may cause trauma to the conjunctiva.

Do not use if package is open or damaged.

Do not use a single speculum and/or depressor on both eyes.

Single patient use only.



